

October 2, 2017

Peter McLaughlin
City of Westfield - Westfield Water Dept.
450 Westfield Water Resource Dept., 28 Sackett Street
Westfield, MA 01085

Project Location: Westfield, MA
Client Job Number:
Project Number: [none]
Laboratory Work Order Number: 17I0994

Enclosed are results of analyses for samples received by the laboratory on September 21, 2017. If you have any questions concerning this report, please feel free to contact me.

Sincerely,

A handwritten signature in black ink that reads "Emily Snyder". The signature is written in a cursive, flowing style.

Emily E. Snyder
Project Manager

Table of Contents

Sample Summary	3
Case Narrative	4
Sample Results	5
17I0994-01	5
17I0994-02	6
17I0994-03	7
17I0994-04	8
Sample Preparation Information	9
QC Data	10
Miscellaneous Organic Analyses	10
B186878	10
Flag/Qualifier Summary	11
Certifications	12
Chain of Custody/Sample Receipt	13

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

City of Westfield - Westfield Water Dept.
450 Westfield Water Resource Dept., 28 Sackett Str
Westfield, MA 01085
ATTN: Peter McLaughlin

REPORT DATE: 10/2/2017

PURCHASE ORDER NUMBER: 2180509

PROJECT NUMBER: [none]

ANALYTICAL SUMMARY

WORK ORDER NUMBER: 1710994

The results of analyses performed on the following samples submitted to the CON-TEST Analytical Laboratory are found in this report.

PROJECT LOCATION: Westfield, MA

FIELD SAMPLE #	LAB ID:	MATRIX	SAMPLE DESCRIPTION	TEST	SUB LAB
Well #1	1710994-01	Drinking Water		EPA 537	
Well #1 Field Blank	1710994-02	Drinking Water		EPA 537	
Well #2	1710994-03	Drinking Water		EPA 537	
Well #2 Field Blank	1710994-04	Drinking Water		EPA 537	

CASE NARRATIVE SUMMARY

All reported results are within defined laboratory quality control objectives unless listed below or otherwise qualified in this report.

The results of analyses reported only relate to samples submitted to the Con-Test Analytical Laboratory for testing. I certify that the analyses listed above, unless specifically listed as subcontracted, if any, were performed under my direction according to the approved methodologies listed in this document, and that based upon my inquiry of those individuals immediately responsible for obtaining the information, the material contained in this report is, to the best of my knowledge and belief, accurate and complete.

A handwritten signature in black ink, appearing to read "Lisa Worthington", is written over a light gray rectangular background.

Lisa A. Worthington
Project Manager

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

Project Location: Westfield, MA

Sample Description:

Work Order: 1710994

Date Received: 9/21/2017

Field Sample #: Well #1

Sampled: 9/21/2017 10:00

Sample ID: 1710994-01

Sample Matrix: Drinking Water

Miscellaneous Organic Analyses

Analyte	Results	MCL/SMCL			Units	Dilution	Flag/Qual	Method	Date	Date/Time	Analyst
		RL	MA	ORSG					Prepared	Analyzed	
# Perfluorobutanesulfonic acid (PFBS)	3.8	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
# Perfluorohexanoic acid (PFHxA)	7.1	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
# Perfluoroheptanoic acid (PFHpA)	2.4	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
# Perfluorohexanesulfonic acid (PFHxS)	28	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
# Perfluorooctanoic acid (PFOA)	6.8	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
# Perfluorooctanesulfonic acid (PFOS)	14	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
Perfluorononanoic acid (PFNA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
Perfluorodecanoic acid (PFDA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
NMeFOSAA	ND	2.0			ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
Perfluoroundecanoic acid (PFUnA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
NEtFOSAA	ND	2.0			ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
Perfluorododecanoic acid (PFDoA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
Perfluorotridecanoic acid (PFTrDA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
Perfluorotetradecanoic acid (PFTA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:24	BLM	
Surrogates		% Recovery	Recovery Limits								
13C-PFHxA		100	70-130						9/30/17 16:24		
13C-PFDA		106	70-130						9/30/17 16:24		
d5-NEtFOSAA		94.7	70-130						9/30/17 16:24		

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

Project Location: Westfield, MA

Sample Description:

Work Order: 1710994

Date Received: 9/21/2017

Field Sample #: Well #1 Field Blank

Sampled: 9/21/2017 10:00

Sample ID: 1710994-02

Sample Matrix: Drinking Water

Miscellaneous Organic Analyses

Analyte	Results	MCL/SMCL			Units	Dilution	Flag/Qual	Method	Date	Date/Time	Analyst
		RL	MA	ORSG					Prepared	Analyzed	
Perfluorobutanesulfonic acid (PFBS)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Perfluorohexanoic acid (PFHxA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Perfluoroheptanoic acid (PFHpA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Perfluorohexanesulfonic acid (PFHxS)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Perfluorooctanoic acid (PFOA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Perfluorooctanesulfonic acid (PFOS)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Perfluorononanoic acid (PFNA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Perfluorodecanoic acid (PFDA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
NMeFOSAA	ND	2.0			ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Perfluoroundecanoic acid (PFUnA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
NEtFOSAA	ND	2.0			ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Perfluorododecanoic acid (PFDoA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Perfluorotridecanoic acid (PFTrDA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Perfluorotetradecanoic acid (PFTA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:12	BLM	
Surrogates		% Recovery	Recovery Limits								
13C-PFHxA		90.7	70-130						9/30/17 16:12		
13C-PFDA		114	70-130						9/30/17 16:12		
d5-NEtFOSAA		92.3	70-130						9/30/17 16:12		

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

Project Location: Westfield, MA

Sample Description:

Work Order: 1710994

Date Received: 9/21/2017

Field Sample #: Well #2

Sampled: 9/21/2017 11:30

Sample ID: 1710994-03

Sample Matrix: Drinking Water

Miscellaneous Organic Analyses

Analyte	Results	MCL/SMCL			Units	Dilution	Flag/Qual	Method	Date	Date/Time	Analyst
		RL	MA	ORSG					Prepared	Analyzed	
# Perfluorobutanesulfonic acid (PFBS)	4.7	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
# Perfluorohexanoic acid (PFHxA)	8.1	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
# Perfluoroheptanoic acid (PFHpA)	2.6	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
# Perfluorohexanesulfonic acid (PFHxS)	35	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
# Perfluorooctanoic acid (PFOA)	7.3	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
# Perfluorooctanesulfonic acid (PFOS)	14	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
Perfluorononanoic acid (PFNA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
Perfluorodecanoic acid (PFDA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
NMeFOSAA	ND	2.0			ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
Perfluoroundecanoic acid (PFUnA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
NEtFOSAA	ND	2.0			ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
Perfluorododecanoic acid (PFDoA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
Perfluorotridecanoic acid (PFTrDA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
Perfluorotetradecanoic acid (PFTA)	ND	2.0	2		ng/L	1	EPA 537	9/22/17	9/30/17 16:37	BLM	
Surrogates		% Recovery	Recovery Limits								
13C-PFHxA		109	70-130						9/30/17 16:37		
13C-PFDA		117	70-130						9/30/17 16:37		
d5-NEtFOSAA		96.2	70-130						9/30/17 16:37		

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

Project Location: Westfield, MA

Sample Description:

Work Order: 1710994

Date Received: 9/21/2017

Field Sample #: Well #2 Field Blank

Sampled: 9/21/2017 11:30

Sample ID: 1710994-04

Sample Matrix: Drinking Water

Miscellaneous Organic Analyses

Analyte	Results	MCL/SMCL		Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
		RL	MA ORSG							
Perfluorobutanesulfonic acid (PFBS)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Perfluorohexanoic acid (PFHxA)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Perfluoroheptanoic acid (PFHpA)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Perfluorohexanesulfonic acid (PFHxS)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Perfluorooctanoic acid (PFOA)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Perfluorooctanesulfonic acid (PFOS)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Perfluorononanoic acid (PFNA)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Perfluorodecanoic acid (PFDA)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
NMeFOSAA	ND	2.0		ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Perfluoroundecanoic acid (PFUnA)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
NEtFOSAA	ND	2.0		ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Perfluorododecanoic acid (PFDoA)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Perfluorotridecanoic acid (PFTrDA)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Perfluorotetradecanoic acid (PFTA)	ND	2.0	2	ng/L	1		EPA 537	9/22/17	9/30/17 14:46	BLM
Surrogates		% Recovery	Recovery Limits			Flag/Qual				
13C-PFHxA		93.2	70-130						9/30/17 14:46	
13C-PFDA		111	70-130						9/30/17 14:46	
d5-NEtFOSAA		94.6	70-130						9/30/17 14:46	

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

Sample Extraction Data

Prep Method: EPA 537-EPA 537

Lab Number [Field ID]	Batch	Initial [mL]	Final [mL]	Date
17I0994-01 [Well #1]	B186878	250	1.00	09/22/17
17I0994-02 [Well #1 Field Blank]	B186878	250	1.00	09/22/17
17I0994-03 [Well #2]	B186878	250	1.00	09/22/17
17I0994-04 [Well #2 Field Blank]	B186878	250	1.00	09/22/17

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

QUALITY CONTROL

Miscellaneous Organic Analyses - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B186878 - EPA 537										
Blank (B186878-BLK1)										
Prepared: 09/22/17 Analyzed: 09/30/17										
Perfluorobutanesulfonic acid (PFBS)	ND	2.0	ng/L							
Perfluorohexanoic acid (PFHxA)	ND	2.0	ng/L							
Perfluoroheptanoic acid (PFHpA)	ND	2.0	ng/L							
Perfluorohexanesulfonic acid (PFHxS)	ND	2.0	ng/L							
Perfluorooctanoic acid (PFOA)	ND	2.0	ng/L							
Perfluorooctanesulfonic acid (PFOS)	ND	2.0	ng/L							
Perfluorononanoic acid (PFNA)	ND	2.0	ng/L							
Perfluorodecanoic acid (PFDA)	ND	2.0	ng/L							
NMeFOSAA	ND	2.0	ng/L							
Perfluoroundecanoic acid (PFUnA)	ND	2.0	ng/L							
NEtFOSAA	ND	2.0	ng/L							
Perfluorododecanoic acid (PFDoA)	ND	2.0	ng/L							
Perfluorotridecanoic acid (PFTrDA)	ND	2.0	ng/L							
Perfluorotetradecanoic acid (PFTA)	ND	2.0	ng/L							
Surrogate: 13C-PFHxA	37.0		ng/L	40.0		92.6	70-130			
Surrogate: 13C-PFDA	39.8		ng/L	40.0		99.4	70-130			
Surrogate: d5-NEtFOSAA	136		ng/L	160		85.3	70-130			
LCS (B186878-BS1)										
Prepared: 09/22/17 Analyzed: 09/30/17										
Perfluorobutanesulfonic acid (PFBS)	1.73	2.0	ng/L	1.77		97.5	70-130			
Perfluorohexanoic acid (PFHxA)	1.83	2.0	ng/L	2.00		91.3	70-130			
Perfluoroheptanoic acid (PFHpA)	1.85	2.0	ng/L	2.00		92.4	70-130			
Perfluorohexanesulfonic acid (PFHxS)	1.91	2.0	ng/L	1.82		105	70-130			
Perfluorooctanoic acid (PFOA)	2.34	2.0	ng/L	2.00		117	70-130			
Perfluorooctanesulfonic acid (PFOS)	1.46	2.0	ng/L	1.85		78.9	70-130			
Perfluorononanoic acid (PFNA)	2.30	2.0	ng/L	2.00		115	70-130			
Perfluorodecanoic acid (PFDA)	2.44	2.0	ng/L	2.00		122	70-130			
NMeFOSAA	2.26	2.0	ng/L	2.00		113	70-130			
Perfluoroundecanoic acid (PFUnA)	2.34	2.0	ng/L	2.00		117	70-130			
NEtFOSAA	2.04	2.0	ng/L	2.00		102	70-130			
Perfluorododecanoic acid (PFDoA)	2.10	2.0	ng/L	2.00		105	70-130			
Perfluorotridecanoic acid (PFTrDA)	1.98	2.0	ng/L	2.00		99.0	70-130			
Perfluorotetradecanoic acid (PFTA)	2.04	2.0	ng/L	2.00		102	70-130			
Surrogate: 13C-PFHxA	39.5		ng/L	40.0		98.7	70-130			
Surrogate: 13C-PFDA	50.7		ng/L	40.0		127	70-130			
Surrogate: d5-NEtFOSAA	177		ng/L	160		111	70-130			

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

FLAG/QUALIFIER SUMMARY

*	QC result is outside of established limits.
†	Wide recovery limits established for difficult compound.
‡	Wide RPD limits established for difficult compound.
#	Data exceeded client recommended or regulatory level
ND	Not Detected
RL	Reporting Limit
DL	Method Detection Limit
MCL	Maximum Contaminant Level

Percent recoveries and relative percent differences (RPDs) are determined by the software using values in the calculation which have not been rounded.

No results have been blank subtracted unless specified in the case narrative section.

CERTIFICATIONS

Certified Analyses included in this Report

Analyte	Certifications
<i>EPA 537 in Drinking Water</i>	
Perfluorobutanesulfonic acid (PFBS)	VT-DW,ME
Perfluorohexanoic acid (PFHxA)	VT-DW,ME
Perfluoroheptanoic acid (PFHpA)	VT-DW,ME
Perfluorohexanesulfonic acid (PFHxS)	VT-DW,ME
Perfluorooctanoic acid (PFOA)	NH,NY,VT-DW,ME
Perfluorooctanesulfonic acid (PFOS)	NH,NY,VT-DW,ME
Perfluorononanoic acid (PFNA)	VT-DW,ME
Perfluorodecanoic acid (PFDA)	VT-DW,ME
NMeFOSAA	VT-DW
Perfluoroundecanoic acid (PFUnA)	VT-DW,ME
NEtFOSAA	VT-DW
Perfluorododecanoic acid (PFDoA)	VT-DW,ME
Perfluorotridecanoic acid (PFTrDA)	VT-DW,ME
Perfluorotetradecanoic acid (PFTA)	VT-DW,ME

The CON-TEST Environmental Laboratory operates under the following certifications and accreditations:

Code	Description	Number	Expires
AIHA	AIHA-LAP, LLC - ISO17025:2005	100033	02/1/2018
MA	Massachusetts DEP	M-MA100	06/30/2018
CT	Connecticut Department of Public Health	PH-0567	09/30/2019
NY	New York State Department of Health	10899 NELAP	04/1/2018
NH-S	New Hampshire Environmental Lab	2516 NELAP	02/5/2018
RI	Rhode Island Department of Health	LAO00112	12/30/2017
NC	North Carolina Div. of Water Quality	652	12/31/2017
NJ	New Jersey DEP	MA007 NELAP	06/30/2018
FL	Florida Department of Health	E871027 NELAP	06/30/2018
VT	Vermont Department of Health Lead Laboratory	LL015036	07/30/2018
ME	State of Maine	2011028	06/9/2019
VA	Commonwealth of Virginia	460217	12/14/2017
NH-P	New Hampshire Environmental Lab	2557 NELAP	09/6/2018
VT-DW	Vermont Department of Health Drinking Water	VT-255716	06/12/2018
NC-DW	North Carolina Department of Health	25703	07/31/2018

17IO994

http://www.contestlabs.com

Doc # 381 Rev 0 5 8 2015

39 Spruce Street
East Longmeadow, MA 01028

CHAIN OF CUSTODY RECORD

Phone: 413-525-2332
Fax: 413-525-6405
Email: info@contestlabs.com



Westfield Water Resources
28 Sackett Street
1-(413)-569-0668
PFAA Testing
Westfield
P.O. # 2180509
Rob Hartman
Hannah Larkham
RH + MG

Company Name: Westfield Water Resources

Address: 28 Sackett Street

Phone: 1-(413)-569-0668

Project Name: PFAA Testing

Project Location: Westfield

Project Number: P.O. # 2180509

Project Manager: Rob Hartman

Con-Test Bid: Hannah Larkham

Invoice Recipient: RH + MG

Sampled By: RH + MG

Enhanced Data Package Required:

Email To: r.hartman@cityofwestfield.org

Email To: h.miller@cityofwestfield.org

Format: PDF EXCEL

Other:

Requested Turnaround Time

7-Day 10-Day

Other: ASAP

Rush-Approval Required

1-Day 3-Day

2-Day 4-Day

Data Delivery

Format: PDF EXCEL

Other:

Enhanced Data Package Required:

Email To: r.hartman@cityofwestfield.org

Email To: h.miller@cityofwestfield.org

Format: PDF EXCEL

Other:

Requested Turnaround Time

7-Day 10-Day

Other: ASAP

Rush-Approval Required

1-Day 3-Day

2-Day 4-Day

Data Delivery

Format: PDF EXCEL

Other:

Enhanced Data Package Required:

Email To: r.hartman@cityofwestfield.org

Email To: h.miller@cityofwestfield.org

Format: PDF EXCEL

Other:

of Containers
Preservation Code
Container Code

Dissolved Metals Samples
 Field Filtered
 Lab to Filter

Orthophosphate Samples
 Field Filtered
 Lab to Filter

Matrix Codes:
GW = Ground Water
WW = Waste Water
DW = Drinking Water
A = Air
S = Soil/Solid
SL = Sludge
O = Other (please define)

Preservation Codes:
I = Iced
H = HCL
M = Methanol
N = Nitric Acid
S = Sulfuric Acid
B = Sodium Bisulfate
X = Sodium Hydroxide
T = Sodium Thiosulfate
O = Other (please define)

Container Codes:
A = Amber Glass
G = Glass
P = Plastic
ST = Sterile
V = Vial
S = Summa Canister
T = Tedlar Bag
O = Other (please define)

Con-Test Work Order#	Client Sample ID / Description	Beginning Date/Time	Ending Date/Time	Composite	Grab	Matrix Code	Conc Code
1	Well #1	7/21/17 10:30			Grab	DW	C
2	Well #1 Field Blank	7/21/17 10:30			Grab	DW	C
3	Well #2	7/21/17 11:30			Grab	DW	C
4	Well #2 Field Blank	7/21/17 11:30			Grab	DW	C

ANALYSIS REQUESTED	PFCS (EPA 537)
	2
	1
	2
	1

Comments:
Please use the following codes to indicate possible sample concentration within the Conc Code column above:
H - High; M - Medium; L - Low; C - Clean; U - Unknown

Detection Limit Requirements
MA
CT
Others: 2.0 mg/L

Program Information
 MCP Analytical Certification Form Required
 RCP Analytical Certification Form Required
 MA State DW Form Required
PWSID # 1329000

NELAC and AIHA-LAP, LLC Accredited

Relinquished by: (signature) Date/Time: 9/21/17 15:30
Received by: (signature) Date/Time: 9/21/17 15:30
Relinquished by: (signature) Date/Time:
Received by: (signature) Date/Time:
Relinquished by: (signature) Date/Time:
Received by: (signature) Date/Time:

TURNAROUND TIME (BUSINESS DAYS) STARTS AT 9:00 AM THE DAY AFTER SAMPLE RECEIPT UNLESS THERE ARE QUESTIONS ON THIS CHAIN. IF THIS FORM IS NOT FILLED OUT COMPLETELY OR IS INCORRECT, TURNAROUND TIME CANNOT START UNTIL ALL QUESTIONS HAVE BEEN ANSWERED.

PLEASE BE CAREFUL NOT TO CONTAMINATE THIS DOCUMENT

39 Spruce St.
 East Longmeadow, MA. 01028
 P: 413-525-2332
 F: 413-525-6405
 www.contestlabs.com



con-test[®]
 ANALYTICAL LABORATORY

Doc# 277 Rev 5 2017

Login Sample Receipt Checklist - (Rejection Criteria Listing - Using Acceptance Policy) Any False Statement will be brought to the attention of the Client - State True or False

Client Westfield Water Resources
 Received By JM Date 9/2/17 Time 1530
 How were the samples received? In Cooler T No Cooler _____ On Ice T No Ice _____
 Direct from Sampling _____ Ambient _____ Melted Ice _____
 Were samples within Temperature? 2-6°C F By Gun # 1 Actual Temp - 7.5
 By Blank # _____ Actual Temp - _____
 Was Custody Seal Intact? N/A Were Samples Tampered with? N/A
 Was COC Relinquished? T Does Chain Agree With Samples? T
 Are there broken/leaking/loose caps on any samples? F
 Is COC in ink/ Legible? T Were samples received within holding time? T
 Did COC include all pertinent Information? Client T Analysis T Sampler Name T
 Project T ID's T Collection Dates/Times T
 Are Sample labels filled out and legible? T
 Are there Lab to Filters? N/A Who was notified? _____
 Are there Rushes? N/A Who was notified? _____
 Are there Short Holds? N/A Who was notified? _____
 Is there enough Volume? T MS/MSD? N/A
 Is there Headspace where applicable? N/A Is splitting samples required? N/A
 Proper Media/Containers Used? T On COC? N/A
 Were trip blanks received? N/A Acid _____ Base _____
 Do all samples have the proper pH? N/A

Vials	#	Containers:	#	#
Unp-		1 Liter Amb.		1 Liter Plastic
HCL-		500 mL Amb.		500 mL Plastic
Meoh-		250 mL Amb.		250 mL Plastic
Bisulfate-		Col./Bacteria		Flashpoint
DI-		Other Plastic		Other Glass
Thiosulfate-		SOC Kit		Plastic Bag
Sulfuric-		Perchlorate		Ziplock
Frozen:				

Unused Media

Vials	#	Containers:	#	#
Unp-		1 Liter Amb.		1 Liter Plastic
HCL-		500 mL Amb.		500 mL Plastic
Meoh-		250 mL Amb.		250 mL Plastic
Bisulfate-		Col./Bacteria		Flashpoint
DI-		Other Plastic		Other Glass
Thiosulfate-		SOC Kit		Plastic Bag
Sulfuric-		Perchlorate		Ziplock
Frozen:				

Comments: